

UV254 Go! Portable



Issue Date: April 2019

Distributed by:

Lab Unlimited
CARL STUART GROUP

Tallaght Business Park
Whitestown, Dublin 24,
Ireland
D24 RFK3

Tel: (01) 4523432
Fax: (01) 4523967
E-mail: info@labunlimited.com
Web: www.labunlimited.com

Quatro House, Frimley Road,
Camberley,
United Kingdom
GU16 7ER

Tel: 08452 30 40 30
Fax: 08452 30 50 30
E-mail: info@labunlimited.co.uk
Web: www.labunlimited.co.uk

About Us



Company Mission

Photonic Measurements are the experts in UV254. We can provide solutions for the measurement of UV254 in all applications within the drinking and waste water treatment, industrial effluent, receiving water and environmental testing, amongst others. Photonic Measurements range of UV254 products allow for the measurement of UV254 and other parameters calibration from the UV254 signal such as TOC, DOC, COD and BOD.

Established in 2014, Photonic Measurements is located in Lisburn, Northern Ireland, UK.



Contents	Page No
Safety Advice	4
Introduction	5
Start-Up Guide	6
Guidelines for Use	8
Menu Screen	9
Data Screen	10
Measurement Routine	11
Measure Menu	12
Calibration Menu	14
Generate Curve	16
View Curve	19
Entering Equation	19
Calibration Example	20
View Equation	28
Settings Menu	29
Data Export	30
Care and Maintenance	32
Additional Parts	33
Path Length	34
Ranges for Surrogates	35
Specifications	36
Glossary	37
Appendices	38
1. What is Absorption	38
2. SUVA Measurement	38
3. TOC Calibration	39
4. DOC Calibration	43
Contact Us	44

Safety Advice

When using the UV254 Go! please ensure you read the manual thoroughly and follow the instructions provided.

Ensure that all necessary equipment provided in the kit is of the standard described upon purchase.

Do not use this instrument for any other usage aside from what is instructed within the manual.

Please ensure that USB port is clean and dry prior to use, remove any object that could interfere with access to the charging port.

The UV254 Go! does not require any specific cleaning for daily operation. However, a dry cloth can be used to collect dust. Do not use cleaning agents or solvents on the UV254 Go!

Introduction

Photonic Measurements' UV254 Go! is a light-weight, portable device, which fits in the palm of your hand. With rechargeable Li-ion battery applications the UV254 Go! is a lab quality device that provides quick and easy measurement. Simply place a sample in the analyser cuvette and you will get your results within seconds.

- A simple icon-based user interface makes it easy to navigate between measuring, calibration and data collection.
- Long service life.
- Extensive storage capacity to preserve numerous measurements.
- Ability to generate graphical charts of data patterns.
- A USB port to connect and transfer data to your PC or Mac.
- Excel compatible, connect with device to examine data further.
- Ability to display up to four surrogate (Calibrated parameters) measurements simultaneously. For example, TOC (Total Organic Carbon).

Start-Up Guide




UV254 Go! Start Up Guide



1. Sample Chamber
2. Sample Chamber Lid
3. USB Charging Port with Cap
4. USB Charging Connection Socket
5. On/Off switch
6. Product Label and Serial Number
7. Touchscreen Menu


Guidelines for Use

The following steps should be used to ensure the Go! is used correctly and efficiently to obtain accurate results.



ONLY use a clean sample cuvette and rinse before use.

With a dry cloth wipe the cuvette to ensure no fingerprints are present on the sample cuvette




Gently tap the cuvette on a flat surface to remove any bubbles from the sample

Sample cuvettes that have scratch marks or have been contaminated should be safely removed.



Menu Screen

The touchscreen menu allows the user to navigate simply through the four main sections of the analyser.



Allows user to view collection of past measurement data.



Takes a UVT/UVA reading, and various surrogate measurements can also be taken.



Add parameters and create a profile for different water sampling points/locations.



Adjust time/date, power and brightness for power save.

Data Menu

Results collected will be stored on the device. Tens of thousands of results can be collected and stored for the duration of the analyser.

To view a result, select the year – you would like to review, and press Select

Continue by narrowing down your selection by choosing a month within that year.

All results obtained within that month will be displayed, select the day you wish to view. All results from that day will be displayed.

To see a plot of results of a period of time, choose one location and one parameter and then press the graph icon.

A graph will be displayed.

Press the grid icon to return to the table view of results.

Folder	File	Name
		con1.go

Select


File Date	Sample	Time	Parameter	Value	Units
07Mar00	Sample0	00	ADC	2500.0	mg/l
	Sample1	05:12:00	ABC	1.0	%
	Sample2	05:12:00	ABC	2.0	%
	Sample3	05:12:00	ABC	3.0	%
	Sample4	05:12:00	ABC	4.0	%
	Sample5	05:12:00	ABC	5.0	%
	Sample6	05:12:00	ABC	6.0	%
	Sample7	05:12:00	ABC	7.0	%
	Sample8	05:12:00	ABC	8.0	%

File Date	Select Location
07Mar00	All
	D
	A

Location D

File Date	Select Parameter
07Mar00	All
	ADC
	ABC

Parameter ADC



Measurement Routine

UVT/UVA- Is the measurement of Ultraviolet Transmission and Absorbance at 254nm.

SUVA- Provides a measurement in relation to EPA415.3 DOC Input from another device will be required for this method. (Please review the standard)

Locations- These can be acquired through user calibration of measurements of parameters/ surrogates that relate to UVA.


Measure Menu

Step 1:

When clicking onto Measure, select from the list of cuvette paths that you will be testing.

See appendix for information on path length.

Press OK to continue.




The cuvette path selected will appear at the bottom of the screen.

Step 2:

Next, begin by selecting the Measurement Routine


For this example, we will select Location1



Step 3:

Place reference cuvette (deionised water) into the Go!


Put the chamber lid back on after placing your cuvette into the chamber



Step 4:

Place sample into Go! and press measure.

To maintain accuracy, use the same cell that was used to take reference point



UVA/UVT are direct measurements

The list on the right-hand side is the user calibrated surrogate measurements


Calibration Menu

The calibrate menu button allows you to create a profile for a range of parameters known to respond to the changes in absorption at 254nm.

Enables you the ability to store these parameters from different sample points as the response can change from location to location.


For calibration procedure examples, please refer to appendices

Select a location you wish to calibrate; the name of the sample can be changed at a later stage.



User calibrated measurements of parameters/ surrogates that correlate to UV254 absorption. Parameters include; TOC, DOC, COD and BOD amongst others calibrated to UV254

Select a Parameter




At this stage if the parameter you selected is not there this can easily be adjusted by simply selecting the item you want to change.




This will also apply if you would like to modify, edit the location, parameter or units name, simple click on the blue box for each component.

When calibrating, four options will appear on screen like so:




1. Generate Curve

Step 1:



After clicking Generate Curve, select the sample size from the list and then click OK




Step 2:

Prepare a reference sample
Then press Reference



Step 3:

Leave reference sample in the Go! and press Measure



Step 4:

Enter the concentration, for the reference cell it will be 0mg/L

The screenshot shows a dark grey input field with the text 'Enter Value' on the left and a text box containing '0 mg/l' on the right.

Step 5:

The points will appear on the graph and the initial reference point (0.0mg/L)


When you are satisfied with your results from the calibration press Review



NOTE: Number of samples will be down to your own lab procedure.


Step 6:

In review check that the maximum value for the calibration is at the top end of the expected range




Click on the editable max value to change it.

Enter a new value if necessary



Step 7:


Click complete to save the calibration



2. View Curve

View Curve

The following graph will be displayed with the current calibration of the parameter selection.



3. Entering an Equation

Enter Equation

For calibration of parameters where the concentration is not known until a later date such as the measurement of BOD 5-Day. We must measure the water on the Go! and record the values. On the sampling day, before the concentration from the primary device is known.

Once the concentration is discovered from the primary device. They can then be plotted in Excel and the coefficients found for the relationship. Finally, the coefficients can be entered onto the Go! The following steps describe this process.

Example: Calibrate 5-Day BOD Test

To calibrate BOD on the Photonic Measurements, Go! we recommend using BOD five-day test.


Take at least 3 samples of the water and divide them into two.

Set 1: one of each for the Go!

Set 2: one each for the five-day BOD test.

(Three samples are needed to obtain enough statistical data to perform a calibration accurately.)

Ensure the samples are labelled to ensure the measurements from each sample can be correlated correctly.



Measure BOD 5-Day


Perform the measurement of BOD using the procedure from the manufacturer of the five-day test. Once complete write the values into a table like so:

Sample Name	Value mg/L
Sample 1	BOD 5 Day measurement
Sample 2	BOD 5 Day measurement
Sample 3	BOD 5 Day measurement


Measure BOD on Go!

Due to the five days test requirement for BOD you will need to measure the absorption of each of the three samples on the day, and then wait for the result to come in from the BOD five-day test. The reason we measure on the GO! before the five-day test is to reduce any growth of organic material in the sample. Thus, instead of going to calibration we click on measurement to get the absorption. Start the Go!

Select the Measure Screen





Click on UVT/UVA to get an absorption value



Prepare a reference sample of deionised water in a cuvette.

Click Reference



Next, Measure Sample 1


Fill cuvette with sample

It is best to use the same cuvette used to take the Reference.


Place a small amount of Sample 1 into the cell (1/10 off cell height) and gently shake sample in the cuvette and empty.

Lastly, place Sample 1 into the cuvette (3/4 full) and place into the Go!

Click Measure



Write down the UVA Value for Sample 1, as shown below in the table. The value 0.506 is used for example purposes only.



Sample Name	Value ABS
Sample 1	0.506
Sample 2	Enter UVA/Absorption value
Sample 3	Enter UVA/Absorption value

Continue this process for Sample 2 and Sample 3, then record values in the table. Lastly, when the 5-Day BOD test is complete, record each of the values.

Calculate the Calibration using Excel

Ensure that you put '0' for both measurements- this will help with creating a straight line.

	B	C
	ABS	mg/L
	0	0
	0.506	33
	0.581	39
	0.426	28

Begin by opening a new spreadsheet and enter the data points collected from both the Go! and BOD 5-day test.


	B	C
	ABS	mg/L
	0	0
	0.506	33
	0.581	39
	0.426	28

Press Insert on the top menu bar in Excel to display data


A screenshot of the Microsoft Excel interface. The ribbon is green and the 'Insert' tab is selected and circled in red. Below the ribbon, the spreadsheet grid is visible. The first row has columns A, B, and C. Column B is labeled 'ABS' and column C is labeled 'mg/L'. The data points are: Row 2: B=0, C=0; Row 3: B=0.506, C=33; Row 4: B=0.581, C=39. A 'Document Recovery' pane is visible on the left side of the spreadsheet area.

	A	B	C	D	E
1		ABS	mg/L		
2		0	0		
3		0.506	33		
4		0.581	39		


When you press Insert across the menu bar click Scatter Graph in Charts



A graph like this one will appear



Right click one of the plotted points and click on Add Trendline



A panel should appear with the following options, select linear curve

Exponential

Linear

Logarithmic

Polynomial Order 2

Power

Moving Average Period 2

Trendline Name

Automatic Linear (Series1)

Custom

Forecast

Forward 0.0 periods

Backward 0.0 periods

Set Intercept 0.0


Display Equation on chart

Display R-squared value on chart

Ensure to tick:

Set Intercept with 0.0 value

Display Equation on Chart



Input calibration data into the Go!

The value displayed will be the B value of the equation

$Mg/L = A + (B \times UVA) + (C \times UVA)^2$ where A and C are zero.

The equation then forms the relationship between the UVA measurement of the Go! and for example, BOD five-day results

Enter in your equation for example

Set A & C to zero and input the value of B found in excel

Change the Max Value of the measurement.

For example, if all three calibration points fall below what you expect the max range to be, then enter the max range of your water.

Lastly, to ensure the calibration is stored in the system. Press Save.

4. View Equation

View Equation

Your equation will appear in the fields on the screen for the given location and parameter.

Location: Location2 Parameter: TOC

Equation of form: $Value = A + (B \times Absorption) + (C \times (Absorption)^2)$

A = B = C =


Max Value = mg/l



Settings Menu


Settings Menu will allow you to adjust certain screen preferences on the device.

Date and Time, can be found in this section by pressing the date button, simply scroll to find the correct date and press the Save button




Press System to adjust screen brightness and hours before reference warning. Managing the brightness will enhance the battery charge.

'Hours Before Reference Warning' simply means after a set time the system will alert you that a reference should be taken.



Information will provide the user with the analyser's software version number and serial number.




Data Export


The UV254 Go! can be connected via USB cable to a PC or MAC. The Go! will act like a USB flash drive by simply clicking into Documents on your device where it should appear on the left-hand side of the menu.

Settings > Info > USB Data Mode


To transfer data to your device, follow these steps: firstly, press to Settings on the Go!




Press Info located at the bottom of the screen



Press USB Data Mode to begin the connection between the GO! and your PC



To access a specific measurement, select the folder starting with the year and month the measurement was taken.




Access Data *.GO in Excel


Click File > Open > Browse > *.go > Enter

Type in *.go as below and click on file.



File name:



A menu bar will open complete the steps as follows: press Delimited



Then ensure tab, comma and other with a colon is ticked and press Finish



Care and Maintenance

Contact with water will not disrupt the use of the UV254 Go! it will not damage the internal components. However, ensure that the device is dry before use.

The UV254 Go! can be cleaned with a dust cloth as regularly as possible.

Battery and Storage



The internal battery is Lithium Polymer. it is recommended that the UV254 Go! device be stored in temperatures between 5 and 27 degrees. Precautions should be taken when leaving the device in a car during extreme weather conditions.

Warranty and Disposal

The UV254 Go! has a 2-year warranty, subject to use according to the manufacturer's instructions.

It is important to note upon disposal of the UV254 Go! device – the Lithium Polymer battery must be fully discharged before taking the device to your local recycling facility.

Additional parts



1. Charger
2. Sample Container
3. Battery Bank
4. Additional Cuvettes
5. USB cable for export data and charging
6. Bottles for easy dispensing samples
7. Cuvettes

Path Length

Sample Cuvette- should be clean and dry before use. Gently place the Cuvette into the cell chamber on top of the Go! see page 6 for Start Up instructions. Cuvettes should be rinsed with deionised water. If sample cuvette contains frosted sides, ensure that the clear sides are facing the light. This can impair the result if not placed in to the chamber correctly.

Path Lengths- is the distance light travels through a sample within a cuvette. Usually a cuvette will be made of quartz between 0.1mL to 10mL size.

The absorbance of the water is not only proportional to the concentration of the material being absorbed but also the optical path length of the sample (Beers Law)

Therefore, UVA and UVT measurements must be given for a certain path length. The path length of 10mm/1cm is used and the absorption is reported abs/cm. This can be set to other values if applicable such as abs/m

The UV254 systems are equipped to facilitate different optical paths. The measurement of absorption is converted to abs/cm no matter what size the optical path. The larger the optical path the more sensitive the measurement will be, whereas smaller optical paths are better suited to measuring higher UV254 absorption such as found in waste water applications.



Ranges for Surrogates

For the UV254 Go! the standard pathlength is 10mm other path lengths. Using a cuvette adaptor smaller cuvette sizes can be used.

Ranges for the UV254 Go!

Parameter	0.5mm path length in mg/L	1mm path length in mg/L	2mm path length in mg/L	5mm path length in mg/L	10mm path length in mg/L
COD	0-5600	0-2800	0-1400	0-560	0-280
BOD	0-2800	0-1400	0-700	0-280	0-140
TOC	0-2400	0-1200	0-600	0-240	0-120
DOC	0-2000	0-1000	0-500	0-200	0-100

Accuracy for the UV254 Go!

Parameter	0.5mm path length in mg/L	1mm path length in mg/L	2mm path length in mg/L	5mm path length in mg/L	10mm path length in mg/L
COD	2	2	1	0.4	0.2
BOD	1	1	0.5	0.2	0.1
TOC	1	1	0.5	0.2	0.1
DOC	1	1	0.5	0.2	0.1

Specification

Measurements	UVA, UVT and SUVA Surrogate measurements TOC, BOD, COD and others
Range	0-100% UVT 0-2.5 ABS, is extendable with shorter pathlength cuvettes
Accuracy	±0.5% UVT
Repeatability	±0.05% UVT
Cuvettes	10mm square UV Quartz Cuvette 1,2,5 & 10mm options
Measurement Time	10 Seconds
Power	Internal Lithium Polymer Battery USB Charger
Wavelength	254nm LED and narrow filter photodiode
Light Source	Deep UV LED Long life, self monitoring
Data Logger	Life time measurement storage.
Dimensions	150x180x77mm
Display	800 x 480 pixel 4in Capacitive Touch TFT colour LCD
Operating Conditions	10 to 45 °C, max 80% relative humidity (non-condensing)
Storage Conditions	-20 to 60 °C, max 80% relative humidity (non-condensing)
Enclosure Rating	IP65
Interfaces	USB for data retrieval
Warranty	2 years

Glossary

Total Organic Carbon (TOC)

(DOC + Suspended Organic Carbon = TOC)

TOC can come from a range of sources such as decaying natural organic material (NOM) and synthetic sources for example industrial chemicals. NOMs can react with chloride and as a result can combine to cause harmful by-products

Dissolved Organic Carbon (DOC)

Part of TOC found in the water. It is the part that combines with chlorine to cause harmful by-products. DOC acts as a food source for bacterial and other microorganisms.

Biochemical Oxygen Demand (BOD)

A measurement of the amount of dissolved oxygen that is needed for aerobic biological organisms in a volume of water to breakdown the amount of organic material at a given temperature over a certain time period. This can be tested using the BOD 5-day test. For water being returned to the environment, the level of BOD is important to maintain aquatic life. Too low- there will be a lack of oxygen and too high- can result in suffocation of aquatic life.

Chemical Oxygen Demand (COD)

The amount of oxygen consumed over a given volume of given in mg/L. COD is different in that COD is the organic compounds that can be chemically oxidised.

Ultraviolet Transmission (UVT)

UVT is the measurement of UV energy at a wavelength or frequency which is transmitted through water

Ultraviolet Absorption (UVA)

UVA is the measurement of light that is absorbed by the sample.

Specific Ultraviolet Absorption (SUVA)

Provides a measurement in relation to EPA415.3 DOC Input from another device will be required for this method.

Surrogate

Is a substitute measurement.

Appendices

Appendix 1: What is Absorption

Absorbance is a measure of the amount of light that is absorbed by the sample.

The absorbance **value A = 2 - log – (T)**

Appendix 2: SUVA Measurement

Environmental agencies such as the EPA (method 415.3) use specific UV absorbance (SUVA) as a measurement for determining the disinfection by-product (DBP). The measurement is the ratio of absorption of UV254 to the dissolved organic carbon (DOC) concentration. The larger the SUVA the more of the total organic material is made up of aromatic organics. Aromatic organics are highly reactive with disinfectants and thus level of aromatics will greatly increase the risk of DBPs.

Appendix 3: TOC Calibration


Tools Needed:

- Photonic Measurement UV254 GO!
- Quartz Cuvette
- Deionised water
- Spectrophotometer or Photometer
- TOC Reagents
- TOC Reagent procedure requirements such as a Block heater.

How to Calibrate TOC:

Firstly, it is recommended using TOC reagents and Photometer from manufactures such as Lovibond and Hach.


Take at least 3 samples of the water and split into two. Like so:




Perform the measurement of TOC using the procedure from the TOC reagent manufacture. Enter the results for each sample in a table for example:

Sample Name	Value mg/L
Sample 1	Enter Photometer measurement here
Sample 2	Enter Photometer measurement here
Sample 3	Enter Photometer measurement here

Start the Go! by using the switch on the back of the device. Press **Calibrate**




Select a location where you will store the calibration on for example **Location1**




Each Location can save up to 4 calibrations. In total 16 calibrations can be saved on the Go!

Select a Parameter for example TOC



On the top of the screen you can change the location, parameter or units name. Simply click on the blue box and enter your updates.

Once complete, select Generate Curve



Prepare a reference sample of deionised water into a cuvette

Then, click the Reference button




Keep the deionised water reference sample in the GO! and then click Measure



Enter the value in the box (mg/L) for the sample.

The reference sample will remain at 0mg/L

When the value has been entered press Measure



Next, measure sample 1.


Reuse the same cuvette used to take reference. Place a small amount of sample 1 into the cuvette, 1/10 off cuvette height and then gently shake the contents within the cuvette and empty

Continue the same process with 3/4 off cuvette height and place into the Go!



Click measure and then enter the value for sample 1. For example, if sample 1 was 65mg/L from the photometer enter it in the box show below.

Continue this process for sample 2 and sample 3.



Three points will appear on the graph along with the point at 0.0mg/L from the reference.


Once all calibration points have been measured click on Review



Edit the Max Value of the measurement. For example, if the three calibration points fall below what you expect the Max Range to be, then enter the Max Range of your water.

For example, 400mg/L

Lastly, press Complete to store your calibration.



Appendix 4: DOC Calibration

NOTE: Follow a similar procedure to that of TOC in the previous appendix

Appendix 5: How to use the UV254 Go! Analyser

It is important to note, that buttons on screen will be coloured green or blue like so:




Blue represents – Active Button

Green represents – Guided Button option (suggested choice)

Operating Touchscreen User Interface:

Touchscreen allows for simple movement between different sections, scroll with finger to navigate through data like so:

File Date	Sample	Time	Parameter	Value	Units
07Mar00	Sample0	00	ADC	2500.0	mg/l
	Sample1	05:12:00	ABC	1.0	%
	Sample2	05:12:00	ABC	2.0	%
	Sample3	05:12:00	ABC	3.0	%
	Sample4	05:12:00	ABC	4.0	%
Location	Sample5	05:12:00	ABC	5.0	%
All	Sample6	05:12:00	ABC	6.0	%
Parameter	Sample7	05:12:00	ABC	7.0	%
All	Sample8	05:12:00	ABC	8.0	%

Scroll up and down

Contact Us


For more information, please contact us at:

3 Crescent Business Park

Lisburn

BT28 2GN

(+44) 02892106263



Distributed by:

Lab Unlimited
CARL STUART GROUP

Tallaght Business Park
Whitestown, Dublin 24,
Ireland
D24 RFK3

Tel: (01) 4523432
Fax: (01) 4523967
E-mail: info@labunlimited.com
Web: www.labunlimited.com

Quatro House, Frimley Road,
Camberley,
United Kingdom
GU16 7ER

Tel: 08452 30 40 30
Fax: 08452 30 50 30
E-mail: info@labunlimited.co.uk
Web: www.labunlimited.co.uk